

To ensure that the test results generated by the DPWH Construction Materials Testing Laboratories (CMTLs) are accurate and reliable, the guidelines for the calibration of testing equipment/apparatus is hereby amended.

As a general policy, the use of calibration, laboratory and field testing equipment/apparatus with no or expired calibration certificate shall be **strictly prohibited.**

All calibration, laboratory and field testing equipment/apparatus of the Bureau of Research and Standards (BRS), whichever is applicable and whenever possible, shall be calibrated by the Department of Science and Technology (DOST) – National Metrology Division (NMD), an ISO 17025-accredited third-party service provider, or by the original manufacturer of the equipment, to ensure that measurements made are traceable to national standards of measurements as specified by the International Committee of Weights and Measures.

The calibration, laboratory and field testing equipment/apparatus of DPWH Regional and Unified Project Management Offices (UPMOs) CMTLs shall be calibrated by the BRS. Meanwhile, the laboratory and field testing equipment/apparatus of DPWH District CMTLs shall be calibrated by their respective DPWH Regional CMTLs. Accordingly, DPWH Regional CMTLs who cannot provide calibration services for DPWH District CMTLs under their jurisdiction shall redirect the request for calibration to the BRS.

Calibration of DPWH UPMOs, Regional, and District CMTLs shall be undertaken as indicated in their submitted **Calibration Plan** (*Annex K*) or as prescribed by the original equipment manufacturer.

All DPWH District CMTLs shall submit the accomplished Calibration Plans to their respective Regional CMTLs. Subsequently, DPWH UPMOs and Regional CMTLs shall accomplish and submit to the BRS their respective Calibration Plan for the succeeding year by the end of November of current year.

Calibration activities shall be completed by the **end of the first quarter of the year**. The Implementing Office shall be responsible for issuing request for calibration of equipment/apparatus which were not calibrated during the said period due to valid circumstances (e.g. equipment under repair during the said period).

Only in cases wherein the DPWH cannot conduct the calibration of laboratory and field testing equipment/apparatus due to valid reasons or circumstances beyond its control may allow the DPWH UPMOs, Regional and District CMTLs to avail the calibration services of other government agencies, Philippine Accreditation Bureau (PAB)-accredited laboratories, or

DPWH-accredited Private Testing Laboratories within their respective regions. However, the calibration shall be witnessed by authorized representatives of the concerned DPWH calibrating entity.

Calibration of laboratory and field testing equipment shall be conducted periodically as specified under *Annex A* or based on manufacturer's operating instructions, or after occurrence of any mechanical or electrical shock that might have affected the equipment. The calibration procedure outlined in *Annex B* shall be strictly observed while the prescribed format of the Calibration/Verification Report (*Annexes C to I*) shall be adopted.

Equipment/apparatus that are not included in *Annex A* does not require calibration, however, verification of physical and mechanical condition, deformations and critical dimensions must be checked before every use and must be compliant with their respective standard reference and/or testing procedure (e.g. ASTM, AASHTO, etc.).

Only qualified personnel with adequate training and experience in calibration works shall be assigned to perform calibration. Likewise, to ensure the integrity of the calibration and for verification purposes, the calibrating entity shall furnish copies of calibration report/certificate of the equipment used for calibration (e.g. Load cells) to the CMTL being calibrated.

Those who seek to attain ISO/IEC 17025 certification may avail calibration services of an accredited third-party provider upon approval of the BRS to satisfy the requirements of such certification.

Compliance monitoring shall be conducted by the BRS to check if the provision under these guidelines are being followed. This may coincide with the spot-checking of Star-Rated DPWH CMTLs pursuant to the provisions of Department Order No. 126, series of 2016. Any subsequent non-compliance hereof may be considered as valid grounds for the downgrading or revocation of Star Rating.

This Order supersedes Department Order No. 53, series of 2019 and shall take effect immediately.

MANUE BONOAN Secretary

14.1.4 RPF/EAA/MLC
Department of Public Works and Highways
Office of the Secretary
WIN2U01797

LIST OF EQUIPMENT FOR CALIBRATION

APPARATUS		MODE OF CALIBRATION	FREQUENCY	REFERENCE STANDARD
1.	Universal Testing Machine (UTM)	Load Cell/Proving Ring	Yearly	ASTM E74/ ASTM E4
2.	Compression Machine/Flexural Machine	Load Cell/Proving Ring	Yearly	ASTM E74/ ASTM E4
3.	Pipe Tester	Load Cell/Proving Ring	Yearly	ASTM E74/ ASTM E4
4.	California Bearing Ratio (CBR) Machine	Load Cell/Proving Ring	Yearly	ASTM E74/ ASTM E4
5.	Marshall Stability Machine	Load Cell/Proving Ring	Yearly	ASTM E74/ ASTM E4
6.	Point Load Tester	Load Cell/Proving Ring	Yearly	ASTM E74/ ASTM E4
7.	Abrasion Machine (Charges/Steel Spheres)	Caliper & Digital Scale	Every 2 Years	ASTM C131/ ASTM C535
8.	Laboratory Oven, Furnace and Other Temperature Measuring Devices	Thermometer	Yearly	ASTM E145
9.	Weighing Scales and Balances	Standard Weights	Yearly	ASTM E617
10.	Set of Sieves	Caliper & Calibration Glass Beads	Yearly	ASTM E11
11.	Moisture Cabinet	Thermo-hygrometer	Yearly	ASTM E145/ ASTM D4230
12.	Gyratory Compactor	Gyratory Compactor Calibration Equipment	Yearly	ASTM D6925
13.	Ductility Apparatus/Machine	Ductility Machine Calibration Equipment	Yearly	ASTM D113
14.	Viscometer	Standard Viscosity Fluids	Yearly	ASTM D445/ ASTM D446
15.	Concrete Air Meter	Concrete Air Meter Brass Calibrator	Yearly	ASTM C231
16.	Caliper	Standard Blocks	Yearly	ASME B89.1.14-2018
17.	Speedy Moisture Tester	Speedy Moisture Tester Calibration Kit	Every 2 Years	ASTM D4944
18.	Rebound Hammer	Anvil	As specified by the Manufacturer	ASTM C805
19.	Load Transducers (Load Cell/Proving Ring)	Universal Calibrating Machine (UCM)	Yearly	ASTM E74/ ASTM E4
20.	Calibration Standard Weights	Higher Class Standard Weights	Yearly	ASTM E617
21.	Calibration Blocks	Calibration Verified Caliper	Yearly	ASME B89.1.14-2018
22.	Universal Calibrating Machine (UCM)	Outsource	As specified by the Manufacturer	ASTM E74/ ASTM E4

CALIBRATION PROCEDURES FOR LABORATORY EQUIPMENT

1. PURPOSE

The purpose of this procedure is to provide a standard calibration procedure, frequency, schedule and requirements for calibration, performance verification, and maintenance of instruments and equipment.

2. SCOPE

This procedure applies to all equipment used in the materials testing laboratories of Bureau of Research and Standards (BRS), Unified Project Management Office (UPMO), Regional and District Engineering Offices.

3. DEFINITIONS

- **Calibration** Adjustment or standardization of the accuracy of the measuring instrument/equipment, usually by comparison with a certified reference or standard.
- **Equipment** Machines or tools used for testing construction materials incorporated in the implementation of DPWH infrastructure projects.
- **Frequency** The schedule of calibration of testing equipment.

4. FREQUENCY

Refer to Annex A.

5. EQUIPMENT

As defined in the Standard Operating Procedures for calibration of equipment. The specific equipment are listed in Annex A.

6. PROCEDURE

- 6.1. Operate the equipment in accordance with the relevant Operation Instructions.
- 6.2. Perform the equipment calibration according to the equipment specific Standard Operating Procedure.
- 6.3. Fill out form "Calibration/Verification Report" Sheet, see Annex C to I.
- 6.4. Attach "Calibration Sticker" to the equipment, see Annex I to K.
- 6.5. The performance verification of the equipment (only where necessary) shall be repeated prior to routine laboratory use.
- 6.6. When external calibrations are performed, service providers that demonstrate competence, measurement capability, and traceability shall be used. Calibration certificates from the providers shall contain the measurement results, including the measurement uncertainty and/or a statement of compliance

7. RE-CALIBRATION

- 7.1. If the calibration does not conform, repeat the procedure.
- 7.2. If the second calibration does not conform, clean and check the equipment according to the manufacturer's instructions. Then, repeat the procedure.

8. LIMITS

Calibration Instructions must be applied in accordance with the equipment specific Standard Operating Procedure. The general limit of accuracy and sensitivity will depend on the manufacturer's manual/instructions.

9. DOCUMENTATION

Document calibration results as per the attached "Calibration/Verification Report" Sheet, see Annex C - I. The Calibration records shall include:

- Name of the laboratory performing the calibration
- Identity of the item of equipment
- Name of manufacturer and model
- Serial/Property Number of the Instrument calibrated
- Laboratory Number of unique Identifier (Calibration Sticker)
- Capacity, Loading range and/or sensitivity
- Date of calibration
- Current location of the equipment
- Manufacturer's instructions or a reference to location
- Reference standard, certified reference material or reference material used for calibration
- Temperature at which the calibration was referenced, if applicable
- Copies of all reports, results of calibration, and/or certificates of calibration
- Maintenance plan and due date for the next calibration
- Identity of the individual performing calibration

ANNEX "C"

Date:	

Lab. Report No.:

CALIBRATION / VERIFICATION REPORT FOR UTM, COMPRESSION/FLEXURAL TESTING MACHINE

Name of Office:	Date of Calibration:
Office Address:	Valid Until:
Equipment/Apparatus:	Capacity:
Brand / Model:	Serial Number:
Brand / Model:	Serial Number:

Applied Force		Readin	g of Loadce	ell (Div)	Average Readin	Error (96)	
TONS	KN	Trial 1	Trial 2	Trial 3	Div	KN	

Remarks:

1. Calibratio	n/verif	ication v	was per	formed u	sing Cor	ntrol	s Digim	ax Plus,	Model		(Serial	No.:]) and a Lo	ad Cell
with	_KN ca	apacity	(Serial	No.:)	in	accord	ance to	ASTM	E74	Standard	Practices	for	Calibratio	n and
Verification	for Fo	rce-Mea	suring	Instrume	<i>nts</i> and	AS'	TM E4	Standa	rd Prace	tices f	for Force I	/erification	of	Testing Ma	achine.
Load cell cal	libratio	n valid u	until:												

2. If the percentage error falls within the allowable error of $\pm 1\%$, (insert recommendation).

CONDUCTED:		REVIEWED:
NAME Designation	NAME Designation	NAME Designation
CHECKED:		ATTESTED:
	NAME Designation	NAME Designation Office

ANNEX "D"

Date

Lab. Report No.:

CALIBRATION / VERIFICATION REPORT FOR CBR/MARSHALL STABILITY MACHINE

Name of Client:	Date of Calibration:	
Office Address:	Valid Until:	
Equipment/Apparatus:	Capacity:	
Brand / Model:	Serial/Property Number:	

Machine Indicated		Reading of I	_oadcell (Div	Calibration Factor			
Load, Div	Trial 1	Trial 2	Trial 3	Average	kN/Division	Lbs/Div	Average
							lbs/div
							1557 417
							Kg/Div

Remarks:

1. Calibration/verification was performed using Controls Digimax Plus, Model ______ (Serial No.: ______), and a Load Cell with ______KN capacity (Serial No.: ______) and calibration valid until ______, in accordance to **ASTM E74** *Standard Practices for Calibration and Verification for Force-Measuring Instruments*.

CONDUCTED:		REVIEWED:			
NAME Designation	NAME Designation		NAME Designation		
CHECKED:		ATTESTED:			
NA	ME		NAME		
Desig	nation		Designation		
			Unice		

ANNEX "E"

Date:

Lab. Report No.:

CALIBRATION / VERIFICATION REPORT FOR ABRASION SPHERES

Name of Client:	Date of Calibration:
Office Address:	Valid Until:
Equipment/Apparatus:	Serial/Property Number:
Brand Model:	

	STEEL SPHERES							
Charges (No. of	Diameter in M	lillimeter (mm)	Weight in Grams (g)					
Spneres)	Actual	Standard (approx.)	Actual	Standard	for ASTM C1	31-14		
#1		46.80		Grading	# of Spheres	Weight:		
#2		46.80		А	12	5000±25 g		
#3		46.80		В	11	4584±25 g		
#4		46.80		С	8	3330±20 g		
#5		46.80		D	6	2500±15 g		
#6		46.80		Average mass of individual: 416 g				
#7		46.80		Standard	for ASTM C5	35-16		
#8		46.80		The charge shall con	nsist of 12-steel	spheres averaging		
#9		46.80		mass between 390 and	approximately 47 mm (1-27/32 m.) in diameter, each na mass between 390 and 445 g, and having a total m 5000±25 g.			
#10		46.80		5000±25 g.				
#11	#11		46.80		Note: Steel spheres or ball bearings 46.0 mm (1-13/16 in.) and			
#12		46.80	and 440 g. each, respect	and 440 g. each, respectively, are readily available. Steel spheres				
Ave. size in mm:		Total Weight in Grams:		or ball bearings 46.8 mm (1-27/32 in.) in diameter having a pproximately 420 g. may also be obtainable.				

Remarks:

1. Calibration/verifaction was performed by using the standard procedure based on **ASTM C131** - *Standard Test Method for Resistance to Degradation of Small-Size Coarse Aggregate by Abrasion and Impact in the Los Angeles Machine* and **ASTM C535** - *Standard Test Method for Resistance to Degradation of Large-Size Coarse Aggregate by Abrasion and Impact in the Los Angeles Machine.*

CONDUCTED: NAME	NAME	REVIEWED: NAME
Designation	Designation	Designation
CHECKED:		ATTESTED:
	NAME Designation	NAME Designation Office

ANNEX "F"

Date:

Lab. Report No.:

CALIBRATION / VERIFICATION REPORT FOR LABORATORY OVEN

Name of Client:	Date of Calibration:	
Office Address:	Valid Until:	
Equipment/Apparatus:	Capacity:	
Brand / Model:	Serial/Property Number:	

Oven Indicated Reading	Thermometer Reading	Temperature Error	Thermostat Reading	IMPORTANT:
°C	°C	°C	Correction (\pm)	If the thermometer reading is within ±5 $^{\circ}$ C
				of the thermostat reading, the calibration
				procedure is completed. If the
				thermometer reading differs from the
				thermostat reading by more than 5 °C, or
				If a more stringent calibration is desired,
				ASIM E145 is to be consulted.

Remarks:

1. Calibration/verification was performed with the use of ______ with serial number ______ and a capacity of ___ °C in compliance with **ASTM E145** - *Standard Specification for Gravity-Convection and Forced-Ventilation Ovens.*

CONDUCTED:		REVIEWED:
NAME Designation	NAME Designation	NAME Designation
CHECKED:		ATTESTED:
N Des	IAME ignation	NAME Designation Office

ANNEX "G"

Date

Lab. Report No.:

CALIBRATION / VERIFICATION REPORT FOR LABORATORY SCALES/BALANCES

Name of Client:	Date of Calibration:	
Office Address:	Valid Until:	
Equipment/Apparatus:	Capacity:	
Brand / Model:	Serial/Property Number:	

Standard / Test Weights	Actual Reading	Error		Max. Allow.
Gram (g)	Gram (g)	Gram (g) Percent		Tolerance, g

Remarks:

1. Calibration/verification was performed with the use of a set of ______ standard weights as per **ASTM E617** - *Standard Specification for Laboratory Weights and Precision Mass Standards.*

CONDUCTED):		REVIEWED:
	NAME Designation	NAME Designation	NAME Designation
CHECKED:			ATTESTED:
	NAME Designation		NAME Designation Office

ANNEX "H"

Date:

Lab. Report No.:

CALIBRATION / VERIFICATION REPORT FOR CALIPER

Name of Client:	Date of Calibration:	
Office Address:	Valid Until:	
Equipment/Apparatus:	Capacity:	
Brand / Model:	Serial/Property Number:	

Standard Length	Measured Reading	Error		Max. Allow.
(mm)	Length (mm)	Millimeter (mm) Percent		Tolerance, mm

Remarks:

1. Calibration/verification was performed with the use of a set of Gauge Block as per **ASME B89.1.14-2018** Standard - Calipers.

CONDUCTE	D:		REVIEWED:
	NAME Designation	NAME Designation	NAME Designation
CHECKED:			ATTESTED:
	NAME Designation		NAME Designation Office

ANNEX "I"

	Lab. Report No.:								
			CA	LIBRA	TION REP	ORT			
Name of Office: Office Address:							Calibration Da Calibration Du	ate: Je:	
Equipment/Apparatus:							Capacity: Serial Numbe	r:	
Reference:	e Standard				Tomporaturo Sta	ndard			
Serial No '	e Stanuaru				Serial No ·	nuaru			
NIST No :					Made:				
Calibrated Date:					Calibration Date:				
Calibration Due:					Calibration Due:				
<i>Specifications of</i> Proving Ring Brand:	the Universal Calibra	ating Machine (UCM):		Indicator Brand:				
Model:					Model:				
Capacity:					Serial No.:				
Serial No.:									
UCM CALIBRATIO	ON FACTOR (lbf/div)	:			Calibration Date: Calibration Due:				
	The interpretat	ion of the Moreh	ouse's issued o	alibration certif	icate can be found on	the issued Calibrat	ion Certificate N	lo•	
APPLI	ED FORCE	READING	G OF LOADCI	ELL (DIV)	AVERAGE READING OF LOADCELL	LOAD PER DIVISION	ACTUAL READING	ERROR (%)	ABOVE ±1%?
KN	DIVISION	TRIAL 1 (0°)	TRIAL 2 (60°)	TRIAL 3 (300°)	DIV	KN/DIV	KN		Recommend
		1	1	1					

Remarks:

1. Calibration/verification was in accordance with ASTM E74 Standard Practices for Calibration and Verification for Force-Measuring Instruments.

2. The calibration was performed using Digimax Plus Digital Display (Brand: Controls; Model: ______) with Serial No.: _____

3. Based on the calibration results, the percentage errors fall within the testing machine tolerable error of +/- 1%.

4. The load cell is recommended for use to calibrate UTM, compression/flexural machine, CBR machine and other force-measuring instruments

CONDUCTED:		REVIEWED:	
(Position)	(Position)	-	(Position)
CHECKED:		ATTESTED:	
	(Position)	_	(Position)

Factor (KN/div.):

Date:

ANNEX "J"



SIZE: 40 X 75 mm

ANNEX "K"

OFFICE LETTERHEAD

Summary Report of Calibrated Equipment and Calibration Plan

Date of Submission

Itom		Voor	Model/		Previous Calibration		Next	Next Calibration	
No.	Laboratory Equipment	Acquired	Manufacturer	Serial Number	Date of last Calibration	Conducted by	Scheduled Date	Assigned Personnel	Condition
1.	XXXXX	2000	XXXXX	000000	01/20/2020	Juan Dela Cruz/BRS	01/20/2021	Juan Dela Cruz/BRS	Operational

Important: Request for calibration shall be submitted at least three (3) months before the expiration date.

PREPARED:

CHECKED:

APPROVED:

NAME Designation NAME Designation NAME Designation